

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130196</u> ✓
Name		<u>William Rapson</u>	ID# <u>6208</u> Date <u>11/8/08</u> ✓
A	Agency	<u>Wasilla Police Dept.</u>	Phone # <u>952-5417</u>
Instrument Location		<u>Wasilla PD</u> ✓	
Alco S/N		<u>X172967</u> ✓	Target Value <u>0.079</u> ✓ High Pressure <u>1250psi</u> ✓
B	Alco Test Values		
		<u>0.083</u> ✓	<u>0.084</u> ✓
		1 st Alco	2 nd Alco
Signature		<u>William Rapson</u> ✓	
(OVER)			

Under the Alaska Rules of Evidence, I certify that (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Rapson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

12/2/08

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

NOVEMBER 03, 2000

OPERATOR'S NAME:

RAFSON, WILLIAM C

OPERATOR'S NUMBER: 6200

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI:

VOC

O.L. #: 123456789

DEPT/AGENCY: WASI

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: X172967

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.78 in

ALCO TARGET .078 15:19

BLANK TEST .000 15:20

INTERVAL STANDARD VERIFIED 15:20

ALCO TV 29.78 in .003 15:21

BLANK TEST .000 15:21

SUBJECT SAMPLE .000 15:22

BLANK TEST .000 15:23

ALCO TV 29.80 in .004 15:23

BLANK TEST .000 15:24

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

NOVEMBER 03, 2000

TIME 15:17

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 45c

BREATH TUBE: 41c

BAROMETER: 29.78 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF

HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop

qrstuvwxyz{|}~